



THESE DRAWINGS/SPECIFICATIONS ARE THE EXCLUSIVE PROPERTY OF MICRO-MEASUREMENTS. THEY SHALL NOT BE REPRODUCED OR COMMUNICATED TO OTHERS, IN WHOLE OR IN PART, NOR BE USED AS A BASIS FOR MANUFACTURE OR SALE OF THE ITEMS REPRESENTED, WITHOUT THE WRITTEN PERMISSION OF MICRO-MEASUREMENTS.

LTR	REVISION	ERN	INITIAL	DATE
A	NOTE 2 - MODEL NUMBER WAS DTSPU15A6-21V	4257	RMB	6-9-10
B	ADDED NOTE #5	5735	RMB	10-3-16

IEC-320 RECEPTACLE





5 PIN DIN MALE PLUG  
 PIN 1 - COMMON  
 PIN 2 - EARTH GROUND  
 PIN 3 - +21 VDC  
 PIN 4 & 5 - NC  
 BARREL - SHIELD/EARTH GROUND



NOTES:

- VENDOR - WALL INDUSTRIES  
5 WATSON BROOK RD.  
EXETER, NH 03833
- MODEL DTGPSU15A6-21V DESK TOP CASE WITH IEC-320 INPUT RECEPTACLE AND 6' [1.8M] OUTPUT SHIELDED CABLE WITH 5 PIN DIN MALE PLUG.  
RATINGS: AC INPUT - 100-240 VAC, 50/60 Hz NOMINAL  
DC OUTPUT - +21VDC, 5%, 0.71A
- DESIGN, MATERIALS & CONSTRUCTION MUST MEET CE LVD REQUIREMENTS.
- VENDOR TO PROVIDE THE CURRENTLY REQUIRED CE LVD RATINGS & CE MARK ON THE CASE.
- CHECK OUT UPON RECEIPT PER PROCEDURE #610-QI0145 (CURRENT REVISION).

ITEM	QUANT	PART NUMBER	DESCRIPTION				
~							
			DRAWN RMB	DATE 4-11-05	 951 WENDELL BLVD. WENDELL, NC 27591 	POWER SUPPLY, DESK TOP 21 VDC, 5%, 0.71A OUTPUT MODEL 700	
			CHECKED	DATE			
MM120-001225	700		PROJ. ENGR JEF	DATE 4-13-05			
NEXT ASSY	USED ON						
MATERIAL		SURFACE TREATMENT		UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES		SCALE	SHEET 1 OF 1
				FRAC	DEC	ANGLES	REV. B
				±1/64	±.005	±1/2°	MM100-122961